

Lea las instrucciones de uso detenidamente antes de empezar a usar el producto.  
Conserve estas instrucciones porque puede tener que volver a leerlas.

## Sibokit

### Test de aliento para el diagnóstico del sobrecrecimiento bacteriano

#### 1. Datos del fabricante

##### Nombre y Domicilio social

ISOMED PHARMA, S.L.

c/París 4, Parque Empresarial Európolis

28232 Las Rozas (Madrid) España

[www.isomed.com](http://www.isomed.com)

#### 2. Contenido del estuche

El envase de SIBO-KIT contiene:

- 1 sobre de lactitol monohidrato (10 g).
- 12 tubos de recogida de muestra de aliento.
- 2 tubos flexibles.
- 1 folleto de instrucciones de uso.
- Etiquetas identificativas de la muestra.

#### 3. Tipo de producto

SIBO-KIT es un test de diagnóstico in vitro de sobrecrecimiento bacteriano. El producto se utiliza para la toma de muestras de aliento de los pacientes que presentan síntomas asociados al sobrecrecimiento bacteriano.

#### 4. Conservación de SIBO-KIT

Ambiente seco y fresco.

#### 5. Instrucciones especiales de utilización

Es importante seguir estas instrucciones de manera precisa para garantizar la fiabilidad y reproducibilidad del diagnóstico. Se recomienda realizar la prueba en presencia de personal cualificado.

#### PREPARACIÓN DEL PACIENTE PREVIA A LA REALIZACIÓN DE LA PRUEBA

- No se recomienda la realización de la prueba en las 4 semanas siguientes a un tratamiento con antibióticos, laxantes osmóticos y/o enemas de limpieza, o a pacientes que hayan sido sometidos a una prueba endoscópica o radiológica que requiera limpieza intestinal.
- Se recomienda la interrupción de la ingesta de probióticos 2 semanas antes de la prueba.
- Debe realizar una dieta pobre en carbohidratos fermentables durante las 24 horas previas al estudio para lograr unos niveles basales de H<sub>2</sub> y/o CH<sub>4</sub> óptimos.
- El paciente tiene que estar en ayuno un mínimo de 12 horas.
- El día de la prueba el paciente realizará la limpieza bucal solo mediante enjuague con una solución de clorhexidina sin sorbitol.
- El paciente debe abstenerse de masticar chicle, fumar y de realizar cualquier esfuerzo físico desde al menos 30 minutos antes del inicio de la prueba y hasta su finalización.
- Durante la realización de la prueba sólo se permite tomar agua sin gas.

## DESARROLLO DE LA PRUEBA

1. Los tubos de recogida de aliento están identificados con el tiempo de toma de muestra tal como se indica en la siguiente tabla:

Nº tubo	1	2	3	4	5	6
Tiempo	BASAL (0 min)	POST (15 min)	POST (30 min)	POST (45 min)	POST (60 min)	POST (75 min)

Nº tubo	7	8	9	10	11	12
Tiempo	POST (90 min)	POST (105 min)	POST (120 min)	POST (135 min)	POST (150 min)	POST (165 min)

2. Antes de la administración del sustrato, el paciente debe **soplar por un tubo flexible** en el fondo del **tubo 1** para obtener la muestra de aliento **BASAL**.

\*Se recomienda que se contenga el aliento durante aproximadamente 30 segundos para que la concentración de H<sub>2</sub>, CH<sub>4</sub> y CO<sub>2</sub> en el aliento sea máxima. La muestra debe contener la última parte del aire espirado. El tubo de vidrio debe taparse rápidamente mientras se continúa soplando con el fin de evitar pérdidas de la muestra.

3. A continuación, se administra el sustrato Lactitol. La dosis recomendada es de 0,5 g/kg de peso hasta un máximo de 10g, disueltos en un vaso con 250 ml de agua si el paciente es adulto. En pacientes pediátricos, se disolverá el sustrato en 125 ml de agua. La prueba también puede realizarse utilizando como sustrato Glucosa (1g/kg de peso hasta un máximo de 100 g) o Lactulosa (0,5 g/kg de peso hasta un máximo de 10g) disueltos en un vaso con 250 ml de agua. Identificar la muestra pegando en la parte frontal de la caja la etiqueta correspondiente al sustrato utilizado que se incluye en el Kit.
4. Tras la administración del sustrato, se recogerán las muestras de aliento cada 15 minutos en los **tubos POST** correspondientes, según la tabla anterior. El procedimiento para la recogida de muestra será igual al descrito en el punto 2.
5. Identificar el kit pegando la etiqueta con el símbolo ① en la zona correspondiente del estuche. La etiqueta con la palabra "**PACIENTE**" debe ser coservada por el paciente, y la identificada con la palabra "**CLIENTE**" debe ser conservada por el centro como comprobante.
6. Enviar el kit junto con la **TABLA DE SÍNTOMAS DEL PACIENTE** al Servicio de Análisis de **ISOMED PHARMA S.L.**

### PROTOCOLO PARA LA REALIZACIÓN DE LA PRUEBA

Se recomienda realizar la prueba en presencia de personal cualificado

## Protocolo SIBO-KIT

1



Recoger la muestra **pre-dosis** (BASAL).

2



Tomar el sustrato disuelto en 250/125 ml de agua.

3



Esperar **15 minutos**.

4



Recoger una muestra **post-dosis** cada **15 min**, durante 165 min.

5



**Identificación** de muestras y **envío** para análisis.

## 6. Incompatibilidades

Dadas las características del producto, no procede.

## 7. Advertencias

Este producto contiene **Lactitol**. Los pacientes con intolerancia hereditaria a la fructosa o a la galactosa, galactosemia o malabsorción de glucosa o galactosa, obstrucción intestinal o desequilibrio de electrolitos en suero no deben utilizar este producto.

Si por cualquier motivo necesita repetir la prueba y volver a tomar el sustrato Lactitol, deben transcurrir al menos 24 h desde la primera administración.

Este producto sanitario **es de un solo uso**. Si se utiliza el producto para más de un paciente, puede producirse contaminación cruzada, haciendo inservible la prueba.

Los tubos de recogida de muestras deben cerrarse correctamente sin ejercer excesiva presión. Se recomienda enroscar hasta que el tapón no gire fácilmente y después girar con cuidado un cuarto de vuelta más.

## 8. Comunicación de incidentes

Si experimenta cualquier tipo de incidente adverso, consulte con un profesional cualificado. También puede comunicarlo directamente al fabricante a través del siguiente correo electrónico: [farmacovigilancia@isomed.com](mailto:farmacovigilancia@isomed.com)



## INSTRUCCIONES PARA EL ENVÍO DE LAS MUESTRAS DE ALIENTO PARA SU ANÁLISIS

- 1.- Rellenar los siguientes datos de la etiqueta identificada con el símbolo “①” que encontrará en el interior del estuche:
  - Código origen
  - Edad
  - Sexo (marcar con una X)
  - Cantidad de sustrato
- 2.- Una vez cumplimentada la etiqueta, pegarla en el espacio del estuche con el símbolo “①” reservado para tal fin.
- 3.- Guardar el resto de etiquetas que identifican la muestra (“CLIENTE/PACIENTE”) como comprobante hasta la recepción del resultado.
- 4.- Enviar el kit (estuche con los tubos tras la recogida del aliento, instrucciones de uso con la tabla de síntomas del paciente cumplimentada y la etiqueta identificada con el símbolo “①” pegada en la zona correspondiente del estuche), a la dirección siguiente:

Isomed Pharma, S.L.  
Avenida Alberto Alcocer, nº 46A, 4ºB  
28016 Madrid  
Email: [analisis@isomed.com](mailto:analisis@isomed.com)

**Es muy importante que en cada muestra indiquen los siguientes datos del remitente para el posterior envío del resultado.**

- **Domicilio**
- **Dirección de correo electrónico (si tiene)**
- **Teléfono de contacto**
- **Relación de las muestras enviadas**

El siguiente formulario se ha establecido para ayudar al profesional sanitario a valorar el resultado del Test de Aliento y poder diagnosticar, con la sintomatología del paciente, si existe sobrecrecimiento bacteriano.

**IMPORTANTE:** Una vez cumplimentado por el paciente, enviar este formulario dentro del kit al Servicio de Análisis de ISOMED PHARMA S.L.

**DATOS DEL PACIENTE**

Código de origen: \_\_\_\_\_

Sustrato administrado: \_\_\_\_\_

Cantidad de sustrato: \_\_\_\_\_

Edad: \_\_\_\_\_

Sexo:

Hombre

Mujer

**VALORACIÓN DE LOS SÍNTOMAS DEL PACIENTE**

Marque con una "X" los síntomas que pudiera padecer durante la prueba.

SÍNTOMAS	ANTES (Ayuno previo)	DURANTE (Toma de muestra)	DESPUÉS (Fin prueba)
Dolor abdominal o cólicos			
Gases (eructos o flatulencias)			
Borborigmos o sonidos abdominales			
Náuseas o vómitos			
Estreñimiento			
Diarrea			
Inflamación o distensión abdominal			
Ardor o acidez estomacal			
Dispepsia o trastornos de la digestión			
Otros síntomas (especificar):			

OBSERVACIONES:

Read the instructions for use carefully before starting to use the product.  
Keep these instructions, as you may have to read them again.

# Sibokit

## Breath test for the diagnosis of bacterial overgrowth

### 1. Manufacturer's information

Registered corporate name and address

ISOMED PHARMA, S.L.

c/París 4, Parque Empresarial Európolis

28232 Las Rozas (Madrid) España

[www.isomed.com](http://www.isomed.com)

### 2. Case contents

The SIBO-KIT package contains:

- 1 sachet of lactitol monohydrate (10 g)
- 12 breath sample collection tubes.
- 2 flexible tubes.
- 1 instruction for use leaflet.
- Sample identification labels.

### 3. Product type

SIBO-KIT is an in vitro diagnostic test for bacterial overgrowth. The product is used for taking breath samples from patients who have symptoms associated with bacterial overgrowth.

### 4. SIBO-KIT conservation

Store in a dry and cool place.

### 5. Special instructions for use

It is important that these instructions be followed precisely to ensure the reliability and reproducibility of the diagnosis. It is recommended that the test be performed in the presence of qualified personnel.

#### **PREPARING THE PATIENT BEFORE THE TEST**

- The conduct of this test is not recommended within 4 weeks of treatment with antibiotics, osmotic laxatives, and/or cleansing enemas, or for patients who have undergone endoscopic or radiological testing requiring bowel cleansing.
- It is advised that the consumption of probiotics be discontinued 2 weeks before the test.
- A diet low in fermentable carbohydrates should be consumed for 24 hours prior to the test to achieve optimal baseline levels of H<sub>2</sub> and/or CH<sub>4</sub>.
- The patient must fast for a minimum period of 12 hours.
- On the day of the test, the patient shall wash their mouth exclusively by rinsing it with a sorbitol-free chlorhexidine solution.
- The patient should refrain from chewing gum, smoking, or performing any physical effort starting at least 30 minutes before the start of the test and until its completion.
- Only non-carbonated water can be consumed during the test.

## TEST PROCEDURE

1. The breath collection tubes must be labeled with the sampling times as shown in the following table:

Tube no.	1	2	3	4	5	6
Time	BASELINE (0 min)	POST (15 min)	POST (30 min)	POST (45 min)	POST (60 min)	POST (75 min)

Tube no.	7	8	9	10	11	12
Time	POST (90 min)	POST (105 min)	POST (120 min)	POST (135 min)	POST (150 min)	POST (165 min)

2. Before administering the substrate, the patient must **blow through a flexible tube** located at the bottom of **tube 1** in order to obtain the **BASELINE** breath sample.  
\*It is recommended that the patient hold their breath for approximately 30 seconds so that the concentration of H<sub>2</sub>, CH<sub>4</sub>, and CO<sub>2</sub> in their breath is maximized. The sample must contain the last portion of the exhaled air. The glass tube should be capped quickly while the patient continues to blow in order to avoid sample losses.
3. The lactitol substrate should be then administered. For adult patients, the recommended dose is 0.5 g/kg of weight (up to a maximum dose of 10 g) dissolved in a glass containing 250 ml of water. For pediatric patients, the substrate must be dissolved in 125 ml of water. The test can also be performed using glucose at a dose of 1g/kg of weight (up to a maximum dose of 100 g) or lactulose at a dose of 0.5 g/kg of weight (up to a maximum dose of 10 g) as substrates, both dissolved in a glass with 250 ml water. Identify the sample by sticking the label on the front of the box included in the Kit corresponding to the substrate used.
4. After administering the substrate, breath samples should be collected into the appropriate **POST tubes** every 15 minutes according to the table above. The sampling procedure shall be the same as described in point 2.
5. Identify the kit by sticking the label with symbol ① to the corresponding area of the case. The label with the word "**PATIENT**" must be kept by the patient, and the one identified with the word "**CUSTOMER**" must be kept by the site as proof.
6. Send the kit together with the **PATIENT'S SYMPTOMS TABLE** to the Analysis Service of **ISOMED PHARMA S.L.**

### TEST PROTOCOL

It is recommended that the test be performed in the presence of qualified personnel

## SIBO-KIT Protocol

1



Collect the **pre-dosing** sample (BASELINE).

2



Take the substrate dissolved in 250/125 ml of water.

3



Wait **15 minutes**.

4



Collect a **post-dosing** sample every **15 min**, for a total of 165 min.

5



Identification of samples and **shipping** for analysis.

## 6. Incompatibilities

Not applicable due to the product's characteristics.

## 7. Warnings

This product contains **lactitol**. Patients with hereditary fructose or galactose intolerance, galactosemia, glucose-galactose malabsorption, intestinal obstruction, or serum electrolyte imbalance should not use this product.

If for any reason you need to repeat the test, please note that at least 24 hours must elapse between the first and second administration of the lactitol substrate.

This medical device is **for single use only**. Cross-contamination may occur if the product is used for more than one patient, thus rendering the test useless.

The sample collection tubes must be closed properly without exerting excessive pressure. It is recommended that the cap be screwed on until it no longer turns easily, and to then carefully turn it another quarter turn.

## 8. Reporting of incidents

If the patient experiences any type of adverse incident, please contact a qualified professional. You may also report any incidents directly to the manufacturer through the following email: [farmacovigilancia@isomed.com](mailto:farmacovigilancia@isomed.com)



### INSTRUCTIONS FOR SENDING BREATH SAMPLES FOR ANALYSIS

1.- Fill in the following data from the label identified by symbol "Ⓢ" that you will find inside the case:

- Source code
- Age
- Sex (please mark with an X)
- Amount of substrate

2.- Once the label has been filled in, stick it to the area of the case with symbol "Ⓢ" reserved for this purpose.

3.- Save the rest of the labels identifying the sample ("CUSTOMER/PATIENT") as proof until receipt of the result.

4.- Send the kit (case containing the tubes after the breath collection, instructions for use with the patient's symptoms table filled in, and label identified by symbol "Ⓢ" attached to the corresponding area of the case) to the following address:

Isomed Pharma, S.L.  
Avenida Alberto Alcocer, nº 46A, 4ºB  
28016 Madrid  
Email: [analisis@isomed.com](mailto:analisis@isomed.com)

**It is very important that the following sender's data be indicated in each sample for the purpose of subsequently resending the result.**

- Address
- Email address (if any)
- Contact telephone number
- List of samples sent

The following form was prepared to help health professionals assess the result of the Breath Test and diagnose bacterial overgrowth based on a patient's symptoms.

**IMPORTANT NOTE:** Once the form has been completed by the patient, please insert it into the kit and send it to the Analysis Service of ISOMED PHARMA S.L.

**PATIENT'S INFORMATION**

Source code: \_\_\_\_\_

Substrate administered: \_\_\_\_\_

Amount of substrate: \_\_\_\_\_

Age: \_\_\_\_\_

Sex: |  Male |  Female

**ASSESSMENT OF THE PATIENT'S SYMPTOMS**

Mark the symptoms that the patient experienced during the test with an "X".

SYMPTOMS	BEFORE (Previous fast)	DURING (Sampling)	AFTER (Test end)
Abdominal pain or cramping			
Gases (belching or flatulence)			
Bowel or abdominal sounds			
Nausea or vomiting			
Constipation			
Diarrhea			
Abdominal inflammation or swelling			
Heartburn			
Dyspepsia or digestive disorders			
Other symptoms (please specify):			

COMMENTS: